

Kreatech[™] FISH probes Product Information Sheet

KBI-40109 Prader-Willi SNRPN (15q11) / PML(15q24)







PI-KBI-40109_D1.1

Kreatech Biotechnology B.V. Vlierweg 20 1032 LG Amsterdam The Netherlands www.LeicaBiosystems.com

Published March 2015



KBI-40109

Kreatech[™] Prader-Willi SNRPN (15q11) / PML (15q24) FISH probe

| Introduction: | Prader-Willi syndrome (PWS) is i include neonatal hypotonia, deve childhood-onset obesity, hypotha approximately 70% of cases this chromosome 15. These deletions SNRPN (small nuclear ribonucle gene region. Angelman syndrom difficult to diagnose approximatel maternally contributed chromoso identified in PWS, but the use of Angelman syndrome. Both syndr detectable by FISH analysis. | a complex, multisystem disor lopmental delay, short statur lamic hypogonadism, and ch is the result of deletion of this s are optimally detected by FI oprotein N, previously known e (AS) is a clinically distinct d y 70% of cases of AS have a me 15. In most cases, this is the more specific probe UBE omes may also result from un | der. The major clinical features e, behavioral abnormalities, aracteristic appearance. In s region from the paternal SH utilizing a probe for the as PWCR, SMN, HCERN3) isorder from PWS that can be deletion of 15q11-q13 in the the same deletion as that 3A is recommended for hiparental disomy which is not | |
|---|---|--|--|--|
| Intended use: | The Prader-Willi SNRPN region probe is optimized to detect copy numbers of the SNRPN gene region at 15q11. The PML (promyelocytic leukemia) gene specific FISH probe at 15q24 is included as control probe. | | | |
| | The probe is recommended to be used in combination with one of the Kreatech Pretreatment kits providing necessary reagents to perform FISH on various sample types for optimal results. (see also <u>www.LeicaBiosystems.com</u> and look for Kits & reagents) | | | |
| Critical region 1 (red): Control region 2 (green): | The Prader-Willi SNRPN specific FISH probe is direct-labeled with Platinum <i>Bright</i> ™550. The PML specific FISH probe is direct-labeled with Platinum <i>Bright</i> ™495. | | | |
| Reagent: | Kreatech probes are direct-labeled DNA probes provided in a ready-to-use format. Apply 10 μl of probe to a sample area of approximately 22 x 22 mm. | | | |
| | Please refer to the Instructions for Use for the entire Kreatech FISH protocol. | | | |
| | Kreatech FISH probes are REPEAT-FREE [™] and therefore do not contain Cot-1 DNA. Hybridization efficiency is increased and background, due to unspecific binding, is highly reduced. | | | |
| Interpretation: | The Prader-Willi SNRPN (15q11) / PML (15q24) FISH probe is designed as a dual-color assay to detect deletions at 15q11. Deletions involving the SNRPN gene region will show one red signal and two green signals at the PML (15q24) control region (1R2G). Two single color red (R) and green (G) signals will identify the normal chromosomes 15 (2R2G) | | | |
| | | Normal Signal Pattern | Del(15q11) SNRPN | |
| | Expected Signals | 2R2G | 1R2G | |
| References: | Mutirangura A et al.Genomics. D Trent RJ et al. J Med Genet. Sep Christian SL et al, Genome Res. | ec;18(3):546-52, 1993 ;34(9):714-8.,1997 Feb;8(2):146-57, 1998 | | |

Warning and precautions: In case of emergencies check SDS sheets for medical advice. SDS sheets may be obtained by either contacting Leica Technical Support or visiting <u>www.LeicaBiosystems.com</u>. DNA probes contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling DNA probes. All materials should be disposed of according to your institution's guidelines for hospital waste disposal.

| Reagent Storage and Handling: | Store at 2-8 °C. Reagents should not be used after the expiration date on the vial label. |
|----------------------------------|--|
| TECHNICAL SUPPORT | Technical support is available at www.LeicaBiosystems.com or +31 20 6919181 or via e-mail: www.LeicaBiosystems.com or +31 20 6919181 or via e-mail: www.LeicaBiosystems.com or +31 20 6919181 |
| CUSTOMER SERVICE | Kreatech probes may be ordered through Leica Customer Service +31 20 6919181 or order via e-mail: purchase.orders@leica-microsystems.com . |