

**PowerVision™ Poly-HRP IHC Detection Systems
(Biotin-free, anti-Mouse and/or Rabbit Primary Antibodies)**

Cat. No.	Alternative Cat. No.	For No. of Slides	Universal IHC Blocking/Diluent	Poly-HRP anti-Mouse IgG and/or anti-Rabbit IgG	Substrate/Chromogen
PV6112	DPVM-15DAB	150	15 mL, RTU	Poly-HRP anti-Mouse IgG (15 mL, RTU)	DAB Substrate Solution (1 mL, 30x) DAB Chromogen Solution (1 mL, 30x)
PV6113	DPVM-110DAB	1100	110 mL, RTU	Poly-HRP anti-Mouse IgG (110 mL, RTU)	DAB Substrate Solution (7 mL, 30x) DAB Chromogen Solution (7 mL, 30x)
PV6114	DPVM-110HRP	1100		Poly-HRP anti-Mouse IgG (110 mL, RTU)	
PV6118	DPVR-15DAB	150	15 mL, RTU	Poly-HRP anti-Rabbit IgG (15 mL, RTU)	DAB Substrate Solution (1 mL, 30x) DAB Chromogen Solution (1 mL, 30x)
PV6119	DPVR-110HRP	1100		Poly-HRP anti-Rabbit IgG (110 mL, RTU)	

I. INTENDED USE

For Research Use Only

PowerVision Poly-HRP IHC Detection Systems are intended for the chromogenic detection of targeted antigens that have been reacted to a user-supplied primary antibody. It is recommended that the reagents are not substituted across detection systems or lot numbers.

II. INTRODUCTION

PowerVision and PowerVision+ IHC Detection Systems utilize a novel poly-labeling technology, wherein secondary antibodies are directly polymerized with HRP or AP into compact polymers bearing a high ratio of enzymes to antibodies. These polymers demonstrated drastically improved detection sensitivity, efficiency and reliability compared to conventional secondary antibody conjugates. Direct polymerization also avoids an endogenous biotin reaction.

III. REAGENTS AND MATERIALS SUPPLIED

For exact catalog number of detection systems and their contents please refer to the above table.

Universal IHC Blocking/Diluent

Ready-to-use, to be used as diluent for primary antibodies or used before primary antibodies for blocking non-specific protein protein interactions.

Poly-HRP anti-Mouse/Rabbit IgG

Ready-to-use, Poly-HRP labeled anti-Mouse IgG and/or anti-Rabbit IgG polymer

Substrate/Chromogen

DAB: Comprises DAB Substrate Solution (30x) and DAB Chromogen Solution (30x).

IV. HANDLING, STORAGE AND SHELF LIFE

Storage Conditions: All reagents are to be stored at 2-8°C. Void after expiration date as specified on detection system/reagent label.

Precautions: Specimens before and after fixation, and all materials exposed to them, should be handled as if capable of transmitting infection and disposed of with proper precautions. Some reagents in this detection system contain hazardous material. The user is advised to consult the MSDS for further information.

V. REAGENTS AND MATERIALS NEEDED BUT NOT SUPPLIED

Universal IHC Blocking/Diluent – for detection systems PV6114 and PV6119.

Substrates/Chromogens – for detection systems PV6114 and PV6119.

Primary antibodies – for all detection systems.

VI. WARNINGS AND PRECAUTIONS

Powervision DAB Chromogen Solution (7mL)

Contains Ethylene Glycol (<90%) and 3,3 Diaminobenzidine Tetrahydrochloride Hydrate (<10%).
GHS07: Exclamation mark.
GHS08: Health hazard.
Signal words: Danger.

H302: Harmful if swallowed.
H341: Suspected of causing genetic defects.
H350: May cause cancer.

P201: Obtain special instructions before use.
P202: Do not handle until all safety precautions have been read and understood.
P264: Wash hands thoroughly after handling.
P270: Do not eat, drink or smoke when using this product.
P281: Use personal protective equipment as required.
P301 + P312: IF SWALLOWED: Call a POISON CENTRE or doctor if you feel unwell.
P330: Rinse mouth.
P308 + P313: IF exposed or concerned: Get medical advice/attention.
P501: Dispose of contents/container to hazardous or special waste collection point.

Powervision DAB Substrate Solution (7mL)

Contains Acetic Acid (<30%).
GHS07: Exclamation mark.
Signal words: Warning.

H315: Causes skin irritation.
H319: Causes serious eye irritation.

P264: Wash hands thoroughly after handling.
P280: Wear protective gloves/protective clothing/eye protection/face protection.
P302 + P352: IF ON SKIN: Wash with plenty of soap and water.
P332 + P313: If skin irritation occurs: Get medical advice/attention.
P305 + P351 + P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
P337 + P313: If eye irritation persists: Get medical advice/attention.
P362: Take off contaminated clothing and wash before reuse.

STAINING PROCEDURE

Each staining run should include both positive and negative tissue control slides to confirm:

1. That the staining system is working properly
2. That positive and negative staining is specific
3. That the correct procedure has been followed.

The combination of antigen retrieval protocol, primary antibody dilution, for use with a detection system should be determined by the user on a series of known positive and negative controls.

The tissue sections should not be allowed to dry out at any point during the rehydration and staining procedures. All procedures are performed at room temperature (18-26 °C).

- 1) Block endogenous peroxidase with 3% hydrogen peroxide in deionised water, for 10 min. Rinse well with wash buffer.
- 2) Block with Universal IHC Blocking/Diluent for 10 min. Blot gently, no need to wash. Note: This step can be omitted if primary antibodies are diluted in the Universal IHC Blocking/Diluent (PV6123)
- 3) Apply primary antibodies for 30-60 min. Rinse well with buffer and soak in buffer for 5 min, twice.
- 4) Apply Poly-HRP anti-Mouse IgG and/or anti-Rabbit IgG and incubate for 30 min. Rinse well with buffer and wash in buffer for 5 min, twice.
- 5) Apply DAB and incubate for 2-5 min. Rinse well with deionized or tap water. Instruction: For DAB: mix one drop (33µl) DAB Substrate Solution & one drop (33µl) DAB Chromogen Solution with 0.93 mL deionized water.
- 6) Counterstain and Mount: Proceed with appropriate counterstaining and mounting protocol.

VII. LIMITATIONS

Correct treatment of tissues prior to fixation and embedding is important for obtaining optimal results. Inconsistent results may be due to variation in fixation, embedding, pre-treatment and primary antibody reactivities, as well as from inherent variations in tissue. Leica Biosystems warrants that the materials sold meet our performance specifications until the expiration, if stored as recommended. No other warranties or guarantees, expressed or implied, are provided, including warranties for merchantability or fitness for a particular purpose.

VIII. GENERAL REFERENCE

- 1) S.R. Shi, J. Guo, R. Cote, L. Young, D. Hawes, Y. Shi, S. Thu and C. Taylor, "Sensitivity and Detection Efficiency of a Novel Biotin-free IHC Detection System: PowerVision", Applied Immunohistochemistry & Molecular Morphology, 7:201-208, 1999
- 2) K. Petrosyan, R. Tamayo, D. Joseph, "Sensitivity of a Novel Biotin-free Detection Reagent (PowerVision+) for IHC" J. Histotechnol, 25:247-250, 2002
- 3) G. Bricca, et al., "Immunostaining Melanoma Frozen Sections: The 1-Hour Protocol" Dermatol. Surgery, 30:403-408, 2004