

**PowerVision™ Poly-AP IHC Reagents
(Biotin-free, anti-Mouse or Rabbit Primary Antibodies)**

Cat. No.	Alternative Cat. No.	For No. of Slides	Poly-AP anti-Mouse IgG or Poly-AP anti-Rabbit IgG
PV6110	DPVM-110AP	1100	Poly-AP anti-Mouse IgG (110 mL, RTU)
PV6133	DPVR-110AP	1100	Poly-AP anti-Rabbit IgG (110 mL, RTU)

I. INTENDED USE

For Research Use Only. Not for use in diagnostic procedures.

PowerVision Poly-AP IHC Reagents are intended for the chromogenic detection of targeted antigens that have been reacted to a user-supplied primary antibody.

II. INTRODUCTION

PowerVision and PowerVision+ IHC Reagents utilize a novel poly-labeling technology, wherein secondary antibodies are directly polymerized with HRP or AP into compact polymers bearing higher ratio of enzymes to antibodies. These polymers demonstrated drastically improved detection sensitivity, efficiency and reliability compared to conventional secondary antibody conjugates. Direct polymerization also avoids an endogenous biotin reaction.

III. REAGENTS AND MATERIALS SUPPLIED

Poly-AP anti-Mouse IgG or Rabbit IgG

Ready-to-use, Poly-AP labeled anti-Mouse IgG or anti-Rabbit IgG polymer

IV. HANDLING, STORAGE AND SHELF LIFE

Storage Conditions:

All reagents are to be stored at 2-8°C. Void after expiration date as specified on reagent label.

Precautions: Specimens before and after fixation, and all materials exposed to them, should be handled as if capable of transmitting infection and disposed of with proper precautions. The user is advised to consult the MSDS for further information.

V. REAGENTS AND MATERIALS NEEDED BUT NOT SUPPLIED

Universal IHC Blocking/Diluent

Substrates/Chromogens

Primary antibodies

VI. WARNINGS AND PRECAUTIONS

**Poly-AP
anti-Mouse IgG
or Poly-AP anti-Rabbit IgG**

Contains 0.009%
2-methylisothiazol-3(2H)-one.

GHS07: Exclamation mark.

Signal words: Warning.

H317: May cause an allergic
skin reaction.

P261: Avoid breathing mist.

P272: Contaminated work clothing should not be allowed out of
the workplace.

P280: Wear protective gloves/protective clothing/eye protection/
face protection.

P302+P352: IF ON SKIN: Wash with plenty of water.

P333+P313: If skin irritation or rash occurs: Get medical attention.

P362+P364: Take off contaminated clothing and wash it before
reuse.

P501: Dispose of contents/container to hazardous or special
waste collection point.

Please note: This reagent contains the biocidal product ProClin™ 950 at 0.09%. The active substance present in ProClin™ 950 is 2-methylisothiazol-3(2H)-one. The ProClin™ 950 biocide is present in this reagent to prevent contamination with microbial organisms.

STAINING PROCEDURE

Each staining run should include both positive and negative tissue control slides to confirm

1. That the staining system is working properly
2. That positive and negative staining is specific
3. That the correct procedure has been followed.

The combination of antigen retrieval protocol and primary antibody dilution for use with a detection system should be determined by the user on a series of known positive and negative controls.

Tissue sections should not be allowed to dry out at any point during the staining procedure. All procedure steps are performed at room temperature (18-26 °C).

1. Block with Universal IHC Blocking/Diluent (PV6123) for 10 min. Blot gently, no need to wash.
Note: This step can be omitted if primary antibodies are diluted in the Universal IHC Blocking/Diluent.
2. Apply primary antibodies for 30-60 min. Rinse well with buffer and wash in buffer for 5 min, twice.
3. Apply Poly-AP anti-Mouse or anti-Rabbit IgG and incubate for 30 min. Rinse well with buffer and soak in buffer for 5 min, twice.
4. Apply AP substrate/chromogen and incubate for 30 min. Rinse well with deionized or tap water.
Note: suitable substrates include Fast Red/Naphthol Phosphate Reagent.
5. Counterstain and Mount: Proceed with appropriate counterstaining and mounting protocol.

VII. LIMITATIONS

Correct treatment of tissues prior to fixation and embedding is important for obtaining optimal results. Inconsistent results may be due to variation in fixation, embedding, pre-treatment and primary antibody reactivities, as well as from inherent variations in tissue. Leica Biosystems warrants that the materials sold meet our performance specifications until the expiration, if stored as recommended. No other warranties or guarantees, expressed or implied, are provided, including warranties for merchantability or fitness for a particular purpose.

VIII. GENERAL REFERENCE

1. S.R. Shi, J. Guo, R. Cote, L. Young, D. Hawes, Y. Shi, S. Thu and C. Taylor, "Sensitivity and Detection Efficiency of a Novel Biotin-free IHC Detection System: PowerVision", Applied Immunohistochemistry & Molecular Morphology, 7:201-208, 1999
2. K. Petrosyan, R. Tamayo, D. Joseph, "Sensitivity of a Novel Biotin-free Detection Reagent (PowerVision+) for IHC" J. Histotechnology, 25:247-250, 2002
3. G. Bricca, et al., "Immunostaining Melanoma Frozen Sections: The 1-Hour Protocol" Dermatologic. Surgery, 30:403-408, 2004