

KI-60004 Tissue Digestion Kit II

For heavily cross-linked paraffin-embedded tissues

RUO - Research Use Only
Not for use in diagnostic procedures

Instructions for use for KI-60004 Tissue Digestion Kit II in combination with Kreatech™ fluorescent labeled DNA probes.

Fluorescent in situ hybridization (FISH) identifies or labels target genomic sequences so that their location can be studied. DNA sequences from appropriate chromosome specific probes are first labeled with reporter molecules. The labeled DNA probe is then hybridized to the metaphase chromosomes or interphase nuclei on a slide. After washing, the specimen is screened for the reporter molecules by fluorescence microscopy.

Kreatech™ FISH probes do not contain Cot-1 DNA. Hybridization efficiency is therefore increased and background, due to unspecific binding, is highly reduced.

This pretreatment kit is specifically developed to obtain optimal results on heavily cross-linked or difficult paraffin embedded tissues (e.g. longer fixation or prolonged storage).

Note: Conventional paraffin embedded tissues
The use of the KI-60007 Tissue Digestion Kit I is recommended for use with conventional samples.

For metaphase and interphase cells or other cytological samples it is recommended to use pretreatment kit KI-60005 or KI-60006.

For more info consult our website: www.LeicaBiosystems.com.

Pretreatment:

Mount 4 - 6 µm formalin-fixed paraffin-embedded tissue sections on positively charged slides (e.g. aminoalkylsilane)

1. Bake mounted slides for 2 hours at 80 °C or 16 hours at 56 °C.
2. De-paraffinize warm slides by soaking in xylene or xylene substitute for two times 10 minutes (min).
3. Re-hydrate by soaking in 100%, 85% and 70% ethanol for 3 min each.
4. Wash with dH₂O for 3 min at room temperature (RT).
5. Pretreat with 0.2 M HCl (LK-099B) for 20 min, and then wash in dH₂O for 3 min at RT.

6. Place slides in Pretreatment Solution B (8% sodium thiocyanate) (LK-100B) at 80 °C for 30 min.
7. Rinse in 2 x SSC (LK-104B) for 3 min at RT.
8. Cover paraffin section with approximately 200 µl Pepsin Solution (LK-101B) and incubate at RT for 5 - 50 min (Time depending on tissue fixation and tissue type. E.g. most breast cancer tissue needs 5 - 15 min digestion; colon tissue 20 - 30 min).
9. Wash in dH₂O for 1 min
10. Wash in 2 x SSC (LK-104B) for 5 min at RT.
11. Dehydrate slides by soaking 70%, 85%, and 100% ethanol for 1 min each time. Air-dry. Proceed with **Probe preparation**:

Note: Check protein digestion and pretreatment by applying 15 µl DAPI counterstain and evaluate slides using a fluorescence microscope equipped with a DAPI filter. Remove cover slip and soak tissue in 2 x SSC for 2 min and prolong protein digestion if sample is not sufficiently digested. Use a fresh sample and reduce protein digestion time if the sample is over-digested.

Probe preparation:

Kreatech™ FISH probes for paraffin embedded tissue are supplied Ready to Use (RtU) unless specified otherwise in the product documentation. Consult label on vial and specific probe pack insert for dilution specifics.

Co-denaturation:

Apply 10 µl of probe per 22 x 22 mm field. Cover with glass cover slip and seal with Fixogum or rubber cement. Denature sample and probe on a Thermobrite (TS-01/02) or on a hotplate at 80 ±1°C for 5 min. Continue with **hybridization**.

Hybridization:

Incubate overnight at 37 ±1 °C in a Thermobrite (TS-01/02) or in a humidified chamber.

Post-Hybridization Wash:

1. Pre-warm Wash Buffer I (0.4 x SSC / 0.3% Igepal) (LK-102) to 72 °C

- Remove rubber cement.
- Place up to 14 slides in 200 ml of Wash Buffer II (2 x SSC / 0.1% Igepal) (LK-103), incubate for 2 min at RT. Slide off coverclips. Re-use only once for a total of 28 slides.
- Place up to 14 slides in 200 ml of pre-warmed Wash Buffer I (0.4 x SSC / 0.3% Igepal) (LK-102), incubate for 2 min at 72 °C (± 1 °C) without agitation. Re-use only once for a total of 28 slides.
- Place up to 14 slides in 200 ml of fresh Wash Buffer II (2 x SSC / 0.1% Igepal) (LK-103), incubate for 1 min at RT without agitation. Re-use only once for a total of 28 slides.
- Dehydrate in fresh 70%, 85% and 100% ethanol, incubate for 1 min each at RT. Air dry at RT and proceed to **counterstaining**.

Counterstaining:

Apply 15 μ l DAPI counterstain (LK-096B) and apply glass cover slip. Proceed with microscopy.

Procedural recommendations:

Temperature and buffer concentration (stringency) of hybridization and washing are important, as lower stringency can result in non-specific binding of the probe to other sequences, and higher stringency can result in a lack of signal. Incomplete denaturation of target DNA can result in lack of signal.

Recommendations for Fluorescence Microscopy:

For optimal visualization use a well maintained and regularly calibrated microscope equipped with a 100 W mercury lamp and a 63x or 100x fluorescent objective. Triple band-pass filters (DAPI/FITC/Cy3 or DAPI/FITC/TRITC) are used to view multiple colors, single band-pass filters are used for individual color visualization.

Suitable excitation and emission range for Kreatech™ fluorophores:

PlatinumBright™415 Ex 429 \pm 20 nm Em 470 \pm 30 nm

PlatinumBright™495 Ex 495 \pm 20 nm Em 525 \pm 30 nm

PlatinumBright™550 Ex 546 \pm 12 nm Em 580 \pm 30 nm

Material provided:

LK-104B	2 x SSC solution
LK-102B	Wash Buffer I
LK-103B	Wash Buffer II
LK-100B	Pretreatment Solution B
LK-099B	0.2 M HCl
LK-101B	Pepsin Solution
LK-096B	DAPI counterstain (1 μ g/ml)
LK-097B	counterstain diluent

Material required, but not supplied:

- Xylene
- Ethanol 100%, 85% and 70%
- Fixogum (LK-071A) or rubber cement
- Hot plate (with accurate temperature control up to 80 °C)
- Incubator with accurate temperature between 37 °C and 56 °C
- Water bath with accurate temperature between 37 °C and 80 °C
- Plastic or glass coplin jars
- Variable micropipettes (1 μ l - 200 μ l)
- Fluorescence microscope equipped with suitable filters (see recommendations for Fluorescence Microscopy).

Warnings and Precautions:

- For in vitro use only. For professional use only.** In case of emergencies check SDS sheets for safety information.
- FISH probes and hybridization buffers contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling FISH probes and DAPI counterstain.
- All materials should be disposed of according to your institution's guidelines for hospital waste disposal.

OSHA HazCom Standard (2012)

Code #	Description	Signal Word	Pictogram	Hazard Statements
LK-096B	DAPI/Antifade 1 μ g/ml (DAPI Counterstain)	Warning		H315, H319, H335
LK-097B	Antifade (Counterstain diluent)	Warning		H315, H319, H335
LK-099B	0.2 M HCl	N.A.	N.A.	N.A.
LK-100B	8% sodium thiocyanate (Pretreatment Solution B)	N.A.	N.A.	N.A.
LK-101B	Pepsin Solution	N.A.	N.A.	N.A.
LK-102B	0.4 x SSC / 0.3% Igepal (Wash Buffer I)	N.A.	N.A.	N.A.
LK-103B	2 x SSC / 0.1% Igepal (Wash Buffer II)	N.A.	N.A.	N.A.
LK-104B	2 x SSC	N.A.	N.A.	N.A.

(*Please note that LK-102 and LK-103 products containing the same number but different letters are identical except for the volume.


A= 100 ml, B=250 ml)

Technical support

Technical support is available at www.LeicaBiosystems.com or toll free at 800-248-0123 or via e-mail: kreatech-support@leicabiosystems.com.

Customer service

Kreatech probes may be ordered through Leica Customer Service toll free at 800-248-0123 or order via e-mail: purchase.orders@leica-microsystems.com.



**Kreatech
Biotechnology B.V.**
Vlierweg 20
1032 LG Amsterdam
The Netherlands

Patents & Tradenames:

Kreatech™ is a trade name of Kreatech Biotechnology B.V., ULS™ is a trademark of Kreatech Biotechnology B.V. Repeat-Free™ is a trademark of Veridex LLC. Thermo-brite™ is a trademark of Statspin.

Nothing disclosed herein is to be construed as recommendation to use any of Kreatech products in violation of any patents. Furthermore, Kreatech does not warrant that the product, either alone, or in combination with other products, is immune from charges of patent infringement. Kreatech will not be held responsible for patent infringements or other violations that may occur from the use of any of its products.